CERTIFICATE OF ANALYSIS

PuriLab Analytical

BLK-RT3-10 10 mg

RT-2026-136 Lot Black Cap Analyzed November 17, 2026



TIMELINE

ORDERED
November 8, 2026

RECEIVED
November 16, 2026

ANALYZED
November 17, 2026

SAMPLE

NAME
BLK-RT3-10 10 mg

EXPECTED
BLK-RT3-10

AMOUNT
1 vial, lyophilized powder

TESTING

CLIENT
Mr. Curious Supplies

TEST TYPE
Content & purity

METHOD
HPLC-UV

CONTENT, HOW MUCH IS IN IT

4.92 mg

Label claim: 10 mg


PURITY, HOW CLEAN IT IS

42.3%

Measured, not estimated. The expected compound was identified at 18.5 min and quantified. A second, larger peak at 19.4 min does not match the expected compound. This sample was submitted as a single compound, not as a blend, and identifying other compounds is outside the scope of this test, so that peak is counted as an impurity in the purity result.

CHROMATOGRAM

HPLC-UV



- 18.51 min, BLK-RT3-10, 42.3%

ANALYST COMMENTS

White lyophilized powder, black-cap vial, received sealed.

ANALYST SIGNATURE

Max Barrett
Lab Director, PuriLab Analytical
Analysis date: November 17, 2026

VERIFICATION CODE

RT-RD7R-SRK4
Verify at purilab.org/verify/RT-RD7R-SRK4

SCAN TO VERIFY

RT-RD7R-SRK4

Results apply solely to the sample as received and are not representative of any larger lot, batch, or production run. Findings are limited to the analytes stated above. PuriLab Analytical is an independent testing laboratory, is not CLIA certified or ISO 17025 accredited, does not sell or distribute any substance, and makes no representation as to the safety, legality, or fitness of any material for any use.